

MRI COMPATIBLE LEAD DESIGNS FOR IMPLANTABLE MEDICAL DEVICES

A THESIS
SUBMITTED TO THE DEPARTMENT OF ELECTRICAL AND
ELECTRONICS ENGINEERING
AND THE INSTITUTE OF ENGINEERING AND SCIENCE
OF BILKENT UNIVERSITY
IN PARTIAL FULFILLMENT OF THE REQUIREMENTS
FOR THE DEGREE OF
MASTER OF SCIENCE

By

Ahmet Ermeydan

September, 2007

I certify that I have read this thesis and that in my opinion it is fully adequate, in scope and in quality, as a thesis for the degree of Master of Science.

Prof. Dr. Ergin Atalar (Supervisor)

I certify that I have read this thesis and that in my opinion it is fully adequate, in scope and in quality, as a thesis for the degree of Master of Science.

Prof. Dr. Ayhan Altıntaş

I certify that I have read this thesis and that in my opinion it is fully adequate, in scope and in quality, as a thesis for the degree of Master of Science.

Prof. Dr. Nevzat G. Gençer

Approved for the Institute of Engineering and Sciences:

Prof. Dr. Mehmet B. Baray
Director of Institute of Engineering and Sciences

ABSTRACT

**MRI COMPATIBLE LEAD DESIGNS
FOR
IMPLANTABLE MEDICAL DEVICES**

Ahmet Ermeydan
M.S. in Electrical and Electronics Engineering
Supervisor: Prof. Dr. Ergin Atalar
September, 2007

It is currently estimated that 600,000 cardiac pacemakers are implanted per year worldwide. It is expected that the usage of other stimulators such as deep brain stimulators (DBS) will reach this number in a short period. On the other hand, 2,000,000 MRI examinations are carried out each year worldwide and usage of MRI is expected to increase.

Unfortunately, people with metallic implants have significant risks in the MRI scanners. It is known that radio frequency and gradient fields of the MRI scanners may induce harmful currents on the implant leads. Radio frequency pulses may cause excessive heating and burns. In addition to this, time-varying gradient magnetic field induced currents on the leads can cause nerve stimulation. In case of cardiac pacemakers, this nerve stimulation may cause cardiac arrest.

In this thesis, novel MRI compatible lead designs were proposed. Lead designs are presented to ensure safe magnetic resonance scanning of patients with active metallic implants such as pacemakers, neurostimulators, and implantable cardio defibrillators. Semiconductor components such as transistors and diodes are used to prevent these undesired induced currents on the implant leads. Circuits on the implants are designed such that while the induction of currents is prevented, the desired signal transmission in between the implanted pulse generator and the body part is maintained.

The designs were tested by using experiments and computer stimulation. It was seen that the new techniques are effective in making MRI safe implantable devices. Benefits and problems of each design will be discussed in this text. It is believed that using this or similar techniques, patients with the implants will be able to be examined safely in MRI scanners.

Keywords: Implantable device, MR Safety, RF heating

ÖZET

VÜCUDA YERLEŞTİRİLEBİLİR MEDİKAL CİHAZLAR İÇİN MRG UYUMLU KABLO TASARIMLARI

Ahmet Ermeýdan
Elektrik-Elektronik Mühendislięi, Yüksek Lisans
Tez Yöneticisi: Prof. Dr. Ergin Atalar
Eylül, 2007

Bugün, dünyada senede 600.000 kalp pili takıldığı tahmin ediliyor. Diğer uyarıcı pillerin kullanımının, örneğin derin beyin uyarıcıları, kısa bir sürede bu rakamı yakalaması bekleniyor. Bununla birlikte, tüm dünya çapında senede 2.000.000 MRG kullanımı gerçekleştirilmekte ve MRG kullanımının daha da artması beklenmektedir.

Ne yazık ki vücuduna metalik cihazlar takılı insanlar MRG cihazları içinde çok ciddi risk altındadırlar. MRG cihazlarının radyo frekansları ve gradyan alanları vücuda yerleştirilen cihazların kablolarında zararlı akımların oluşmasına sebep olabileceği biliniyor. Radyo frekansı sinyallerinden kaynaklanan en önemli risk fazla ısınma ve yanıktır. Buna ek olarak, gradyan alanlar da istenmeyen sinir uyarımlarına sebep olabilir. Kalp pili takılı hastalarda bu uyarım kalbin aniden durmasına sebep olabilir.

Bu tez çalışmasında, MRG uyumlu vücuda yerleştirilebilir kablo tasarımları geliştirdik. Kalp pilleri, siniruyarıcıları, kalp defibrilatörleri gibi aktif metalik vücuda yerleştirilebilir cihazlar taşıyan hastaların güvenli bir şekilde manyetik rezonans görüntülemesi yapılabilmesi için kablo tasarımları yapıldı. Vücuda yerleştirilen cihazların kablolarında meydana gelen bu istenmeyen akımları engellemek için yarı-iletken elemanlar örneğin transistörler ve diyotlar kullanılabilir.

İstenmeyen indüklenme akımını engelleyecek ve darbe üretici ile vücut dokusu arasında sinyal iletimini devam ettirecek şekilde vücuda yerleştirilebilir cihazların devreleri tasarımılandı.

Bu tez içerisinde geliştirilen tasarımlar deneyler ve simülasyonlar ile denendi. Tasarımların MRG uyumlu vücuda yerleştirilebilir cihazlar yapmak için etkili olduğu görüldü. Her bir tasarımın faydaları ve olası problemleri bu yazıda tartışıldı. Yakın bir gelecekte, bu tezde bahsedilen yöntemler veya benzeri tekniklerin kullanılması ile vücuda yerleştirilmiş elektronik uyarıcı taşıyan hastaların güvenli bir şekilde MRG cihazlarında çekimlerinin yapılacaktır.

Anahtar Kelimeler: Vücuda yerleştirilebilir uyarıcılar, MR güvenliği, RF ısınma

ACKNOWLEDGEMENTS

I wish to express my deepest gratitude to my supervisor **Prof. Ergin Atalar**. With his knowledge, he has been a very helpful supervisor. During my M.S. study, I learned a lot about Magnetic Resonance Imaging and fundament of electronics with the guidance of Dr. Atalar. I am very grateful for his support and patience during my M.S. study. In my future life, I will always remember his support and contribution.

I would also like to thank my friends; Yiğitcan Eryaman, Halise Irak, Elif Aydoğdu, Onur Taşcı, Haydar Çelik and professors; Ayhan Altıntaş, Vakur Ertürk, Bülent Özgüler, Nail Akar, Ezhan Karaşan, I would like to thank them for their kind support. I thank Prof. Dr. Nevzat Gencer for taking part in my jury.

I would like to thank to my parents and my wife. Their presence was the biggest motivation for me everyday of this study. My wife, Esra Ermeýdan actively contributed the experiments I've conducted. Without her it would have been impossible.

I would also like to thank TÜBİTAK for financial support during my M.S. study.

TABLE OF CONTENT

1. INTRODUCTION	1
2. BACKGROUND	6
2.1 <u>IMPLANTABLE STIMULATOR</u>	6
2.2 <u>MECHANISM OF HEATING</u>	10
3. DESIGN	13
3.1 <u>ACTIVE LEADS</u>	13
3.2 <u>DIODE RESISTOR CIRCUIT (DRC)</u>	14
3.3 <u>TRANSISTOR DIODE CIRCUIT (TDC)</u>	19
3.4 <u>CAPACITOR SWITCH CIRCUIT (CSC)</u>	24
4. SIMULATIONS	29
4.1 <u>PART I: PACING PULSE MODELS</u>	29
4.1.1 <u>DIODE RESISTOR CIRCUIT (DRC)</u>	29
4.1.2 <u>TRANSISTOR DIODE CIRCUIT (TDC)</u>	30
4.1.3 <u>CAPACITOR SWITCH CIRCUIT (CSC)</u>	35
4.2 <u>PART II: INDUCED VOLTAGE SIGNAL MODEL</u>	37
4.2.1 <u>DIODE RESISTOR CIRCUIT (DRC)</u>	37
4.2.2 <u>TRANSISTOR DIODE CIRCUIT (TDC)</u>	37
4.2.3 <u>CAPACITOR SWITCH CIRCUIT (CSC)</u>	38
4.3 <u>PART III: EFFICIENCY MODEL OF CSC</u>	40
5. EXPERIMENTS	45
5.1 <u>HEATING MEASUREMENT EXPERIMENTS</u>	45
5.1.1 <u>COMPARISON OF WIRE HEATING WITH AND WITHOUT RF BLOCKING ELEMENTS (RFBE)</u>	45
5.1.2 <u>DIODE RESISTOR CIRCUIT (DRC)</u>	48
5.1.3 <u>CAPACITOR SWITCH CIRCUIT (CSC)</u>	50
5.2 <u>NERVE STIMULATION EXPERIMENT</u>	54
5.2.1 <u>CAPACITOR SWITCH CIRCUIT (CSC)</u>	54
6. RESULTS.....	58
6.1 <u>HEAT MEASUREMENT EXPERIMENT RESULTS</u>	58
6.1.1 <u>COMPARISON OF WIRE HEATING WITH AND WITHOUT RF BLOCKING ELEMENTS (RFBE)</u>	58
6.1.2 <u>DIODE RESISTOR CIRCUIT (DRC)</u>	62
6.1.3 <u>CAPACITOR SWITCH CIRCUIT (CSC)</u>	62
6.2 <u>NERVE STIMULATION EXPERIMENT RESULTS</u>	64
6.2.1 <u>CAPACITOR SWITCH CIRCUIT (CSC)</u>	64
7. DISCUSSIONS	65
8. CONCLUSION.....	69
9. REFERENCES.....	70

LIST OF FIGURES & TABLES

Figure 1. Illustration of main structure of the IPG.	6
Figure 2. Illustration of unipolar, bipolar and multielectrode leads.	8
Figure 3. (a) Simplified diagram of stimulator output stage for unipolar pacing mode; (b) Simplified diagram of stimulator output stage for bipolar pacing mode.	9
Figure 4. Simplified diagram of mechanism of electrode heating.	12
Figure 5a. Unipolar pacing mode DRC implementation with capability of programming positive pulses	14
Figure 5b. Unipolar pacing mode DRC implementation with capability of programming positive and negative pulses.	14
Figure 6. Bipolar and unipolar pacing mode DRC implementation with capability of programming positive and negative pulses.	17
Figure 7. Multi electrode pacing mode DRC implementation with capability of programming positive and negative pulses.	18
Figure 8. TDC implementation	19
Figure 9. Bipolar and unipolar pacing mode TDC implementation-1 with capability of programming positive and negative pulses.	21
Figure 10. Multi electrode pacing mode TDC Implementation-1 with capability of programming positive and negative pulses.	22
Figure 11. Unipolar and bipolar pacing mode TDC Implementation-2 with capability of programming positive and negative pulses.	23
Figure 12. Multi electrode pacing mode TDC Implementation-2 with capability of programming positive and negative pulses.	24
Figure 13a. CSC (Parallel Capacitor) implementation	25
Figure 13b. CSC (Series Capacitor) implementation	26
Figure 14. Simulation result of DRC on electrode tissue impedance	30
Table 1. 5Spice PMOS and NMOS Characteristic Data	31
Table 2. Spice Model parameters definitions and units of MOSFETS	33
Figure 15a. Simulation result of TDC for unipolar pacing mode with positive pulse	34
Figure 15b. Simulation result of TDC for unipolar pacing mode with negative pulse	34
Figure 16. Simulation result of TDC for bipolar pacing mode	35
Figure 17a. Simulation result of CSC-parallel capacitor- for bipolar pacing mode	36
Figure 17b. Simulation result of CSC-series capacitor- for bipolar pacing mode	36
Figure 18a. Simulation result of CSC-parallel capacitor- on the electrode tissue impedance at 1 KHz	39
Figure 18b. Simulation result of CSC-parallel capacitor- on the ring tissue impedance at 1KHz	39
Table 3. Simulation result of DRC, TDC and CSC voltage measurement on the electrode and ring impedances.	40
Figure 19. The efficiency plot of CSC as a function of R_1	41
Figure 20. The efficiency plot of CSC as a function of capacitance	42
Figure 21. The efficiency plot of CSC as a function of t_1	43
Figure 22. The efficiency plot of CSC as a function of t_2	43
Figure 23. Semi cylindrical phantom model.	46
Figure 34. Lead placement in the gel	47
Figure 35. Phantom and lead placement in the scanner	47
Figure 26. Lead model without RFB used in the heating measurement	48

Figure 27. Lead model with RFBE used in the heating measurement	48
Figure 28a. Gel phantom setup with DRC	49
Figure 28b. Gel phantom setup with DRC	49
Figure 29a. IPG model without CSC	51
Figure 29b. IPG model with CSC	51
Figure 30a. Rectangular plastic phantom side 1	52
Figure 30b. Rectangular plastic phantom side 2	52
Figure 31a. Experiment setup without CSC	53
Figure 31b. Experiment setup with CSC	53
Figure 32. The Sciatic Nerve bundle where it leaves the vertebral column [27].	54
Figure 33. Sciatic Nerve in thigh musculature [27]	55
Figure 34. Nerve muscle preparation of the frog leg	56
Figure 35. Nerve stimulation experiment setup	57
Figure 36. Temperature rise versus time at the distal and proximal electrode of a 20 cm wire without RF Blocking element	59
Figure 37. Temperature rise versus time at the distal and proximal electrode of the 20 cm wire 3.9 k Ω resistance placed on the distal electrode of the wire	59
Figure 38. Temperature rise versus time at the distal and proximal electrode of a 20 cm wire with PIN diode MA4P7452F-1072T placed on the distal electrode of the wire	60
Figure 39. Temperature rise versus time at the distal and proximal electrode of a 30 cm wire without RF blocking element	60
Figure 40. Temperature rise versus time at the distal and proximal electrode of the 30 cm wire 3.9 k Ω resistance placed on the distal electrode of the wire	61
Figure 41. Temperature rise versus time at the distal and proximal electrode of a 30 cm wire with PIN diode MA4P7452F-1072T placed on the distal electrode of the wire	61
Figure 42. Heating experiment result of IPG model with DRC	62
Figure 49. Heating experiment result of IPG model without CSC	63
Figure 44. Heating experiment result of IPG model with CSC	63

1. INTRODUCTION

With an increasing number of magnetic resonance imaging (MRI) examination and parallel growth in the number of implantable stimulators such as pacemaker, implantable cardioverter defibrillator (ICD) implants and deep brain stimulator (DBS), safety of MRI for the patient with these implants has become an important issue. In this thesis, novel implant designs that enable safe MRI exams are proposed.

More than 370,000 pacemakers and ICDs were implanted in the United States in 2003 [1]. According to Pennsylvania Health Care Cost Containment Council (PHC4), the number of MRI scans increased from 9.3 million to 13.5 million between 1999 and 2001 in USA. PHC4 indicated that 120000 MRI examinations were performed in Pennsylvania hospitals in 2003 [2]. According to these data, it is estimated 50-75% probability of a patient being indicated for an MRI over the lifetime of their device [3]. Magnetic resonance imaging utilizes three types of magnetic fields: static (also known as B_0); radio frequency (RF); and gradient magnetic fields.

The static magnetic field is a critical element in a magnetic resonance imaging system. It is designed to be very strong (0.1 tesla to 7 tesla and typically 1.5 tesla) and very uniform (in the order of 1 parts per million inhomogeneity within volume of interest).

In the MRI scanner, an RF magnetic field with the frequency of 42.58 MHz/T is generated by RF coils with the aim of exciting the spins that are magnetized by static magnetic field.

The RF field is typically circularly polarized and perpendicular to the static magnetic field.

The gradient fields are applied to encode signal generated by the excited spins. They also serve in the slice selection process. In the MRI scanners, there are three gradient fields.

Using the combination of the x-, y- and z-gradient fields, a gradient field in any direction can be obtained enabling slice selection and encoding in any arbitrary direction.

Excessive heating in the presence of elongated metallic structures such as electrical wires, guidewires, and implant electrodes inside the body have been investigated by several researchers. Smith *et al.*[4] measured RF heating inside a 1.5T MRI scanner. In this gel-phantom study, the temperature increase was measured in different lengths of wires with 0.6 mm insulation while an SAR of 4.2 W/kg was applied to the body. The result of this study shows that the maximum temperature rise was measured as 16.82 °C at the end of the 40 cm wire.

Sommer *et al.* [5] conducted a study on twenty-one models of pacemakers and forty-four pacemaker electrodes in order to evaluate safety and feasibility of magnetic resonance imaging at 0.5 T in patients with implanted cardiac pacemakers. *In vitro* measurements of this study show that maximum temperature rise at electrode of the leads was measured as 8.90 °C at SAR of 0.6 W/kg. The temperature increased by 23.50 °C when SAR of 1.3 W/kg was applied. *In vivo* measurements were made with forty-four patients with implanted cardiac pacemakers. After 51 MRI examinations, none of the pacemakers displayed a dysfunction at MRI and no changes occurred in the programmed parameters.

Luechinger *et al.* [6] measured temperature rise in saline phantom at the electrodes for different types of pacing leads. The maximum temperature rise was measured approximately 13 °C for unipolar lead. Besides, *in vivo* experiments were performed in a pig. The maximum temperature rise was measured as 15 °C for a passive fixation lead and it was measured as 30 °C with screw in lead. No significant threshold or impedance changes of the leads were observed.

In one reported case [7], a patient with DBS became comatose after 15 minutes of diathermy. The RF magnetic field is used to heat tissue in this procedure. After this procedure, the patient was unresponsive and comatose condition was confirmed after neurological exams. Edema centered at the electrodes was showed in the MRI examination.

In another case [8], a thermocoagulation lesion adjacent to the electrode after MRI examination was observed on a patient with DBS system. The MRI examination was performed at 1 T with body transmit coil. But the manufacturer (Medtronic, Minneapolis, MN) specifies cautions for the MRI examination of a DBS patient and it is explained that head transmit coil with SAR less than 0.4 W/kg at 1.5 T should be used for DBS patients during MRI examinations.

Rezai *et al.* [9] measured RF heating with gel-phantom. The maximum temperature rise was measured as 25.3 °C with an average SAR of 3.9 W/kg and there is no loop in the lead for this measurement. On the other hand, the leads were positioned with two small loops that are approximately 2.5 cm in diameter in order to simulate a loop near the burr-hole cover for DBS lead placement. The temperature rise was measured as 6.1 °C at this condition.

Kainz *et al.* [10] was performed another study about heating of neurostimulating systems. The measurements were performed at 1.5 T and 3.0 T in the anatomical realistic and cubic phantoms. The temperature rise was measured as 2.1 °C at the electrode of the lead. It was concluded that reduction of the number of the loops at the IPG reduces heating associated in MRI examinations.

Bendadis *et al.* [11] performed experiments with 27 MRI examinations on 25 patients. 26 MRI scans were performed with head transmit coil, while one of them was performed with body transmit coil. One of the patient reported chest pain during MRI examination. This survey yielded no reports of discomfort around the lead or IPG excluding one case.

Dempsey *et al.* [12] investigated mechanism RF heating during MRI examination. It was concluded that resonant circuits with inductance and capacitance have significant heating potential. Thus extended wires also have potential for thermal injury.

There should be limitations to make safe MRI examinations and these criteria are defined by The United States Food and Drug Administration (FDA). These criteria are stated that during

MRI examination, SAR should be in the limit of 8 W/kg in the head and trunk; 12 W/kg in the extremities when averaged over a gram of tissue for 15 minutes. Also, temperature changes should be in the limit of 1 °C in the head, 2 °C in the trunk and 3 °C in the extremities [13].

Several strategies may be used to ensure safety in metallic implants. One of the basic methods is limiting the input RF Power. Determination of RF power threshold provides to specify reasonable amount of heating during MRI examination. Yeung et al. [14] presented a methodology for such a power limitation. However, many modern pulse sequences such as fast spin-echo or steady-state free precession (SSFP) require high RF power levels. Therefore there is no guarantee that good quality images can be acquired with this limitation.

Susil et al [15] presented usage of RF chokes in the design of a combined electrophysiology/MRI catheter. Besides, Ladd M.E et. al. [16] presented usage of triaxial chokes to present high impedance to currents flowing on the outer surface of the triax.

Implantable medical device incorporating integrated circuit notch filters was developed by Medtronic (Minneapolis, MN) [17], and pacemaker with EMI protection was developed by Teletronics [18]. These two designs provide to safety for electromagnetic interference to the implant. However, limitation of RF heating at the lead is not guaranteed. High current may still be flowing through long cables, which can cause excessive heating and burns.

Biophan's "photonic pacemaker" provides optical signal transmission in between the generator and the body part [19]; which provides safety [20]. Because there is no coupling with the optical system and the electromagnetic field, the conversion between electrical and optical systems has limited efficiency and therefore the lifetime of the pulse generator reduces significantly. The miniaturization of this system is also a difficult task.

Another possible safety problem with MRI is that gradient induced currents on the implants may cause undesired nerve stimulation with a possibility of cardiac arrest. A spinal fusion

stimulator was analyzed theoretically for this purpose [21] and although authors have noted a reduction in the stimulation threshold level change was not alarming. As will be discussed in this manuscript this conclusion is not inline with my experimental results.

As it will be explained more detailed in the next section, the active implants are composed of two parts: implantable pulse generator (IPG) and a lead. IPG has typically a metallic case and the lead is insulated with a plastic coating and the electrode of the lead is bare and touches the target organ. In some designs (bipolar leads), two wires are used in each lead. One of these wires is connected as ground and connects to a relatively large metallic ring. In some designs more than two wires are used and connected to a series of electrodes. Even in other designs, multiple leads are used. In this study, a novel method is proposed to make MRI safe active implanted devices. This method uses active circuit elements on the lead in order to make MRI safe lead. This new approach is not only easier to manufacture and miniaturize, but also more effective in reducing the induced currents than passive circuits such as choke inductors.

2. BACKGROUND

In this chapter, the necessary background for explaining the design of the MRI compatible implantable stimulators will be given.

2.1 IMPLANTABLE STIMULATOR

Implantable stimulators are composed of one implantable pulse generator (IPG) and one or more leads. Although IPG has different functionalities for different types of stimulators such as pacemakers, ICDs and DBS, general functionalities can be summarized as: (1) the creation of stimulation pulses on the target tissue; (2) the communication with a programmer which may be placed outside of body; (3) other functions like sensing. Structure of the IPG can be divided into four main parts: (1) electronic circuits; (2) battery; (3) female connector; and (4) case. This structure is illustrated in Figure 1.




Figure 1. Illustration of main structure of the IPG.

In order to perform stimulation and sensing functions of IPG, there should be connection between IPG and target tissue. This connection is provided by the leads. Their mechanical structure varies with stimulator types. According to applied therapy, pacemakers and ICDs have two main types of leads; unipolar and bipolar (see Figure 2). The stimulators operating in unipolar pacing mode use only one wire in the pacing lead. In this method, the pacing pulse is transmitted by applying potential between the pacing lead and the case of the IPG (implantable pulse generator). The stimulators operating in a bipolar pacing mode use two wires in the pacing leads. In this method, one wire is used as the pacing signal and the other one is used as ground. For example, Kappa 900-KDR901 model of Medtronic pacemaker can operate in both bipolar and unipolar pacing modes, while Kappa 900-KDR906 model of Medtronic pacemaker [22] uses only unipolar lead. On the other hand, DBS and pain stimulators utilize multielectrode leads. Multielectrode leads typically have four or more electrodes (see Figure 2). A wide variety of current patterns can be obtained using these electrodes. Since the current applied to each of the electrodes can be modified independently, it is possible to obtain any combination of unipolar and bipolar pacing.




Figure 2. Illustration of unipolar, bipolar and multielectrode leads.

Note: Leads with a wide variety of lengths are provided by the manufacturers, typically in the ranges of 30cm to 80cm.

Modern stimulators have advance functionality in order to deliver programmable pulses. The output stage of a typical stimulator is shown in Figure 3. With the aid of this circuit stimulation voltage can be programmed from 0.2V to around 7.0V with pulse widths ranging from 0.1 ms to 2 ms. The typical stimulator battery voltage is in the range of 2 to 2.8 V. Since the pacing amplitude may be higher than battery voltage a charge pump with a tank capacitor is used in most IPGs. A series capacitor is placed in the circuit in order to null the total charge delivered to the tissue. This effect of nulling the total charge helps to overcome effect of ‘electrode polarization’, which is formed by accumulation of charges around the

electrode-tissue interface [23]. Tank and series capacitors are typically in the order of $10\mu\text{F}$. The typical impedance seen at the electrode and tissue interface is in the range of $400\ \Omega$ to $1000\ \Omega$ [23]. Simplified diagram of stimulator output stage is illustrated in Figure 3.




Figure 3. (a) Simplified diagram of stimulator output stage for unipolar pacing mode; (b) Simplified diagram of stimulator output stage for bipolar pacing mode.

For unipolar pacing, the ground of electronic circuits is connected to stimulator case (Figure 3a). During charging period, S1 and S3 switches are closed and S2 switch is open and tank capacitor is charged by the charge pump.

During stimulation period, S1 and S3 switches are open and S2 switch is closed the accumulated charges on the tank capacitor flows over the series capacitors to the tissue.

When the short stimulation period ends, the switches return back to original conditions, i.e. S1 and S3 switches are closed and S2 switch is open. In this period, the tank capacitor gets lost charges back from the charge pump, and in the mean time the charges accumulated in the series capacitor discharges toward ground. The operating principle of the bipolar pacing is very similar to the unipolar case. Only difference is that in the bipolar pacing, the ground of electronic circuits is connected to the tissue by the lead rather than the case (see Figure 3b).

Modern cardiac pacemakers also have important functionalities. One of the most critical functionality is sensing bioelectric signals from the cardiac muscle (intracardiac ECG). This function provides synchronous pacing and recording any abnormal cardiac activity. Besides, defibrillators also use sensing in order to detect fibrillation. Typical intracardiac ECG signal amplitude ranges from 0.2 mV to 3.2 mV in the atrium and from 0.4 mV to 6.4 mV in the ventricle [23]. The detection of these small signals requires amplification and filtering. The main network power (50 Hz in Europe and 60 Hz in USA) interfere with the intracardiac ECG signal. Other interference source is the electromyogram signal originating from the muscles around the pacemaker case. This electromyogram signal frequency range changes from 100 Hz to a few kHz. Since the energy of intracardiac ECG signal has significant components at low frequencies a 70 to 200 Hz pass-band filtering is applied in order to minimize the interference [23].

2.2 MECHANISM OF HEATING

In order to explain the proposed methods of preventing the excessive temperature rise at the electrode of the implant lead, a very brief explanation of the heating mechanism is given below.

In MRI, radio frequency (RF) pulses are applied to obtain echo from the sample of interest. The frequency of RF pulse is proportional with the strength of the main magnetic field,

which is approximately 64 MHz for a 1.5T MRI scanner. Magnetic fields are only required to obtain the desired echo, however, since it is not possible to uncouple electric and magnetic fields, some additional and unwanted electric field is generated in the body. The optimization process (the process of obtaining the desired magnetic field while minimizing the electric field) yields different results for different type of magnets. For a horizontal bore magnet, the optimum field distribution that minimizes the electric field while keeping the magnetic field uniform in the body is achieved by a special coil type called birdcage coil [24]. In a properly designed birdcage coil, electric field at the center of the object is zero and increases linearly in the radial direction, while electric field is oriented in the z-direction (along the axis of the magnet bore).

When a metallic implant is placed inside the body, the electromagnetic field may induce current on the wire. The worst conditions are when the lead is directed mostly in the z-direction and loops are formed by the lead.

Concentrating on one lead with one electrode case, electric and magnetic field in the body generates a potential difference between pacemaker case and the electrode. This voltage causes a current flow on the lead. This current typically is not high enough to cause heating of the wires of the lead but the current leaves cable at the electrode. High concentration of the current at the tissue adjacent to the electrode heats up significantly. Radio frequency currents may flow from lead directly to the body by the mechanism known as displacement current. In this mechanism, the lead insulation material acts as dielectric and high currents may still flow [4-9].

Calculation of the amount of heating at the tissue adjacent to the electrode is a complex procedure. It involves solution of an electromagnetic scattering problem in a lossy medium and also a bioheat problem in an inhomogeneous medium. Induced voltage on the stimulator experiences impedance which is combination of the electrode-tissue impedance,

lead impedance, lead-to-IPG case impedance, IPG to body impedance. In a typical implant design, these other impedances can be ignored. Induced current is determined by the electrode-tissue impedance (see Figure 4).




Figure 4. Simplified diagram of mechanism of electrode heating.

3. DESIGN

In this chapter, three different MRI-compatible lead designs are given.

3.1 ACTIVE LEADS

Here, the active leads have been proposed with the aims of reducing lead electrode heating problem. Three different implementations have been proposed: “Diode-Resistor Circuit”, “Transistor-Diode Circuit” and “Capacitor-Switch Circuit”.

In the first implementation, PIN-diode and resistor are placed in parallel on the pacing leads in order to reduce the induced current on the leads. Pacing energy is slightly reduced due to the finite turn-on voltage of the diode. However, induced radio-frequency current is reduced significantly since PIN diode acts as a high valued resistor at this frequency. The parallel resistor increases the leakage currents that alleviate the problem of charge accumulation at the electrode. This design can be used in both unipolar and bipolar pacing. Since both diode and resistor are incorporated in this circuit, we call the circuit “Diode-Resistor Circuit” (DRC).

In the second implementation, a resistor is placed in parallel to series combination of transistor and diode. They are placed on to pacing leads to reduce the induced current on the leads at radio frequencies (MRI application) and gradient fields. Similar to first implementation, this circuit enables pacing pulse to be transferred. However, the pacing energy is slightly reduced due to the finite turn-on voltage of the diode and turn-on resistance of the transistor. Induced currents, however, are blocked. This design is called “Transistor-Diode Circuit” (TDC) and can be used both unipolar and bipolar pacing modes.

In the third implementation, resistors are placed on to pacing lead to reduce the induced current on the leads. Pacing energy is accumulated on a serial or parallel capacitor placed at the distal end of the lead and discharged to the target body part by the help of an electronic

switch. Since capacitor is charged with a very low current, impedance of the wire can be made very high. This circuit is called “Capacitor-Switch Circuit” (CSC).

The operating principle of the new approach is much simpler than passive circuits. The circuit can be miniaturized and therefore can be incorporated to flexible leads without a significant effort and cost.

3.2 DIODE RESISTOR CIRCUIT (DRC)

In this implementation, parallel diode and resistor are added on to pacing leads to reduce the induced current on the leads. This implementation is seen in Figure 5.




Figure 5a. Unipolar pacing mode DRC implementation with capability of programming positive pulses




Figure 5b. Unipolar pacing mode DRC implementation with capability of programming positive and negative pulses.

In this implementation, during pacing pulse, diode conducts the current after pacing pulse level passes diode's turn-on voltage threshold level (typically 0.6V). Therefore, there is no significant change in the applied signal level. When the stimulator is exposed to radio frequency electromagnetic radiation, induced RF voltage on the lead would not cause excessive current at the electrode because of the diode and resistor. In the implementation of this design, PIN diode is used. PIN diodes' characteristic is very suitable for this application. Although they behave like a regular diode at a low frequency signal such as pacing signal is applied to them, they act as a resistor at radio frequency (RF). Typically, their RF impedance is high (in the order of several kilo-ohms) when zero or negative bias is applied. Electrical properties of PIN diodes vary and can be obtained from their manufacturers' data sheets.

One of the main characteristics of the Diode-Resistor Circuit is that in the normal operation, i.e., when IPG sends the pacing pulse, it causes only a small loss of power. On the other hand, it exhibits a resistance to radio-frequency signals. This resistance is a function of applied positive current. When no or negative voltage is applied to the diode, the RF resistance is in the order of several kilo ohms. This is very useful property in order to block induced current flow on the lead. In the proposed design, the parallel diode-resistor pair blocks the induced current flow.

Besides, standard stimulators use serial capacitor each lead. There are two significant reasons for the usage of this serial capacitance: (1) it provides safety condition to block DC current flow to the target body part in a fault condition; (2) it eliminates the possibility of electrode polarization due to null the total charge on the body by discharging the capacitor. In order to enable capacitor discharge, a parallel resistor is used in this implementation. The value of this parallel resistor must be chosen properly. If resistance is too high, capacitor

cannot be discharged during one pacing cycle and adversely affect the pacing capability of the design. On the other hand, if the resistance is too low, some RF current may flow on the resistor and therefore the safety performance of the lead design will diminish.

As described in Background section of this thesis, electrode tissue impedance can be selected as 1 k Ω . In the experiments of this thesis, a MACOM PIN Diode MA4P7452F-1072T was used. Its impedance is 3.5 k Ω . If the resistance of the resistor is selected as 10 times larger than impedance of diode, resistance value of the resistor can be ignored in the parallel combination of resistor and diode. Thus, a 50 k Ω resistor is very suitable for this application.

Modern pacemakers have two types of pacing pulse techniques: unipolar and bipolar. The method is selected based on the patient condition and illness. In unipolar pacing mode, pacing signal transmission between electrode and the case of the IPG is performed and this operation is controlled by the IPG.

Above mentioned design works only for positive unipolar pacing pulses. If negative pacing pulses are desired, the orientation of the diode may be reversed. However, if capability of programming both negative and positive unipolar pacing is desired, the design has to be modified as shown in Figure 5b.

In this design, two wires are necessary for each of the electrodes. While one of the wires is in use, the other is disconnected using an electronic switch. Using this method, programmable-polarity unipolar pacing can be obtained, while ensuring high resistance at RF frequencies in order to minimize induced lead currents.

In addition to this, in bipolar pacing mode, two electrodes are used at the same time during pacing sequence and case of the pacemaker is not used. One of the wires carries the signal from voltage source through the PIN diode and it reaches target body part. Then, this pacing pulse returns from body part to ground ring of the lead. This ring connected to IPG through

another wire with PIN diode and pacing pulse reaches the IPG through this wire. In order to create both positive and negative pacing pulses, there should be four wires and four PIN diodes in this implementation. This four-wire and one-electrode design support both unipolar and bipolar pacing modes if the electronic circuit at IPG contains switches that would relay the ground from the ring of the lead to the case of the IPG. Figure 6 shows this pacing configuration employing four independent wires with PIN diodes.




Figure 6. Bipolar and unipolar pacing mode DRC implementation with capability of programming positive and negative pulses.

On the other hand, Deep Brain Stimulators (DBS), Spinal Cord Stimulators (SCS) and Nerve Root Stimulators (NRS) have different stimulation technique rather than common electrical stimulation application such as cardiac pacing and defibrillators. In these simulators, multiple electrodes are used in order to generate complicated current profile in the body with the aim of providing maximum therapeutic benefits [25]. For example, typical DBS generators use four to eight electrodes in today's technology. In these designs, similar to unipolar pacing mode, pulses are programmed independently for each of the electrodes

and the IPG case is used as ground. This capability enables bipolar pacing mode, where the pacing pulse is transmitted between two different electrodes by programming one of the electrode positive and the other electrode negative pulse voltage. The mode of operation and potentials applied to the electrodes are selected by the physician in order to obtain a successful therapeutic effect.

In order to achieve this desired mode of operation, the design has been modified such that each of the electrodes connected to two wires using a PIN diode and resistor pair as shown in Figure 7. Serial capacitors are placed on one of the wires of each electrode. Arbitrarily, these capacitors are placed on the wires that carry positive pulse. Alternatively, the capacitors can be placed on the negative current carrying wires. In order to provide MRI compatibility and the desired functionality of stimulators with multi electrode mode, electrodes of the stimulators are reconfigured as seen in Figure 7.




Figure 7. Multi electrode pacing mode DRC implementation with capability of programming positive and negative pulses.

3.3 TRANSISTOR DIODE CIRCUIT (TDC)

In this implementation, resistor is placed in parallel to series combination of transistor and diode. This design not only eliminates induced RF currents but also minimizes gradient field induced currents. The implementation is in Figure 8.




Figure 8. TDC implementation

Although the transistor can be selected as BJT or MOSFET, in this thesis, the transistor was selected as a MOSFET transistor in order to increase efficiency and decrease power loss. Also enhancement-mode MOSFET is the most suitable model for this design. There are two different options for the diode selection. High frequency diode with high power durability or a PIN diode can be used.

When the pacing pulse is transmitted from IPG to target tissue, IPG opens MOSFET transistor by changing gate voltage of the transistor. At this condition MOSFET transistor is in the triode region and it behaves like a very small resistance. Pacing pulse level is adjusted greater than threshold voltage level of the diode. When the signal voltage level is higher than threshold voltage level of diode, current flows and the target tissue is stimulated.

One of the important advantages of this design over Diode-Resistor Circuit is that TDC does not only blocks induced RF currents but also induced low frequency currents. In this way, possible gradient-induced currents will also be blocked using this circuit.

In order to operate unipolar and bipolar pacing modes, two different implementations are made. In the first implementation, PMOS transistor is used when the positive pacing pulse is transmitted. Based on CMOS technology, very small resistance is obtained. PMOS transistors' source to gate voltage difference should be significantly greater than source to drain voltage difference in order to obtain very small resistance that can be assumed to be short circuit. Thus, during pulse transmission, zero or negative voltages should be applied on the gate of the PMOS transistor and positive voltage should be applied on source of the PMOS transistor. By using this configuration, small resistance value is obtained during transmission of pacing pulse. Diode connection is adjusted based on positive pulse direction with PMOS transistor.

In addition to this, negative pulse transmission requires similar configuration. Therefore, NMOS transistor is used for this purpose. Based on CMOS technology, NMOS transistors' gate to source voltage difference should be significantly greater than drain to source voltage difference in order to obtain very small resistance. Thus, during pulse transmission, required positive voltages should be applied on the gate of the NMOS transistor and zero or negative voltage should be applied on source of the NMOS transistor.

In order to use both unipolar and bipolar pacing modes, electrode of the stimulator is reconfigured and it is seen in Figure 9. However, a handicap exists for this implementation. During transmission of the pacing pulse, one of PMOS transistor should work and the other one should be closed. Similarly, one of NMOS transistor should work depending on selection of PMOS transistor.




Figure 9. Bipolar and unipolar pacing mode TDC implementation-1 with capability of programming positive and negative pulses.

In order to satisfy requirement of multiple functions of DBS, SCS and NRS, TDC implementation should be changed. Similar to Diode Resistor circuit, two wires are required to implement one electrode. In this model, when the one electrode transmits pacing pulse, other electrode cannot transmit any pacing pulse. Thus, two electrodes cannot transmit pulse at the same time. This situation limits some functional requirements but it satisfies many functional applications with ease of implementation. Configuration of multi electrodes lead is seen in Figure 10.




Figure 10. Multi electrode pacing mode TDC Implementation-1 with capability of programming positive and negative pulses.

The second implementation of this model was designed to provide flexibility to first implementation. In the first implementation, each electrode control depends on other electrode. Therefore, gate of the transistors are connected to the IPG and this increases the flexibility. IPG sets the gate voltage but this situation increase number of wires. This configuration is sketched in Figure 11.




Figure 11. Unipolar and bipolar pacing mode TDC Implementation-2 with capability of programming positive and negative pulses.

In multi electrode implementation of this model, each electrode directly controlled by IPG and it means that each electrode can transmit pacing at the same time. This implementation model increases number of wires. This situation can satisfy all functional requirements but it causes usage of more wires. Configuration is seen in Figure 12.




Figure12. Multi electrode pacing mode TDC Implementation-2 with capability of programming positive and negative pulses.

3.4 CAPACITOR SWITCH CIRCUIT (CSC)

In this implementation, when charging the capacitor, the connection between body part and capacitor is kept open. Capacitor is charged with the small current flowing on the resistors in the ideal period between two pace pulses. In order to pace, the accumulated energy on the capacitor is suddenly discharged on the body part. The duration of pacing pulse and the

frequency are programmed to the implantable pace generator by the physician to achieve desired treatment. Typically, while pacing frequency is on the order of one hertz, the discharging period is about 1 msec. This model is suitable for only bipolar pacing. The capacitor can be placed parallel or series to leads and PMOS transistor selected as a switch in this implementation. This implementation models are seen in Figure 13a and 13b.




Figure 13a. CSC (Parallel Capacitor) implementation

In this implementation, there is no additional control line. By using suitable transistor, pacing process is easily adjusted. However, this design may suffer from “electrode polarization” problem since the average current applied to the electrode is not zero.




Figure 13b. CSC (Series Capacitor) implementation

In this implementation, there is a control line in the lead. By applying suitable control voltage on the transistor, pacing process can be easily adjusted. Since there is a series capacitor in the circuit, “electrode polarization” problem does not exist.

In this implementation, capacitor and serial resistors values should be optimized, in order to find most power efficient conditions. Typically PMOS (MOSFET) transistors have very small resistance (in the order of $50\ \Omega$), when transistor is on. Therefore, transistor effect was ignored in the efficiency calculation. During charging, some power dissipates on the serial resistors and some power transmits on the body part. In order to increase the IPG’s battery life, an efficient power transmission needs to be obtained. The power efficiency of this circuit can be obtained. In the following paragraphs, the efficiency equation will be obtained.

After system is used for several seconds, capacitor voltage charge and discharge waveform becomes periodic. We assume period starts at the end of the charging period. We assume this maximum capacitor voltage is $V_{C2}(0)$. When the switch is on, the capacitor discharges over the electrode-tissue interface.

The voltage of the capacitor decreases during discharging period and reaches the value defined as $V_{C1}(0)$. The equation of capacitor voltage during charging can be written as:

$$V_{C2}(0) = V_s + (V_{C1}(0) - V_s) e^{-\frac{t_1}{R_1 C}} \quad (1)$$

where, C is capacitance of capacitor, t_1 is capacitor charging duration (on the order of 1 sec); R_1 is the total series resistance on the capacitor during charging period and V_s is the battery voltage.

Similarly, the equation of capacitor voltage during discharging can be written as:

$$V_{C1}(0) = V_{C2}(0) e^{-\frac{t_2}{R_2 C}} \quad (2)$$

where, C is capacitance of capacitor, t_2 is capacitor discharging duration (on the order of 1 msec); R_2 is the electrode-tissue impedance during discharging period.

During charging period of the capacitor, some power is dissipated on serial resistance (R_1) and equation of this dissipated energy on R_1 can be written as:

$$W_1 = \frac{C}{2} \left[V_s - V_{C2}(0) e^{-\frac{t_2}{R_2 C}} \right]^2 \left(1 - e^{-\frac{2t_1}{R_1 C}} \right) \quad (3)$$

where, W_1 is dissipated energy on serial resistance.

During discharging period of the capacitor, some power is transmitted on the body part (R_2) and equation of transmitted energy to body part can be written as:

$$W_2 = \frac{C}{2} V_{C2}(0)^2 \left(1 - e^{-\frac{2t_2}{R_2 C}} \right) \quad (4)$$

where, W_2 is transmitted energy to body part.

By using these two equations, efficiency equation of this system can be written as:

$$\eta = \frac{W_2}{W_1} = \frac{\left(-1 + e^{\frac{t_1}{R_1 C}}\right)^2 \left(1 - e^{\frac{-2t_2}{R_2 C}}\right)}{\left(\left(1 - e^{\frac{-t_2}{R_2 C}}\right) e^{\frac{t_1}{R_1 C}}\right)^2 \left(1 - e^{\frac{-2t_1}{R_1 C}}\right)} \quad (5)$$

where, η is efficiency of the system.

4. SIMULATIONS

In this chapter, in order to observe whether proposed designs satisfy expectations about pacing pulse transmission and RF current blocking, pacing pulse models and induced voltage models due to MRI simulations of three different MRI-compatible lead designs are done. Pacing pulse simulations, induced voltage simulations of three implementations and efficiency simulation of CSC are given in the following part.

4.1 PART I: PACING PULSE MODELS

Basic properties of simulations should be defined before simulation results will be given. In standard stimulator applications, pulse duration changes with the interval of 0.1ms to 2ms. In this implementation, pulse duration is applied as 1msec with the period of 1 sec. electrode-tissue impedance is assumed to be 1 k Ω . Pulse level is applied as 5 volt.

4.1.1 DIODE RESISTOR CIRCUIT (DRC)

In order to simulate this implementation, OrCAD PSpice 9.1 Demo (Cadence, 2655 Seely Avenue, San Jose, California 95134, USA) was used. In this simulation, Infineon Technologies BA595 pin diode model was used. 1msec pulse was applied in the period of 1 sec on the diode and resistor. As a result of this simulation, 4.2 volt pulse level was observed on the electrode tissue impedance at 1msec. This result is seen in Figure 14. For bipolar pacing mode, pulse level is expected as 3.6 volt on the electrode tissue impedance because of two diode threshold voltage.




Figure 14. Simulation result of DRC on electrode tissue impedance

4.1.2 TRANSISTOR DIODE CIRCUIT (TDC)

In order to simulate this implementation, 5Spice Electronic Circuit Simulation Program (Non-commercial Version, Richard P. Andresen, 2004) was used. In this simulation 1N5711 Schottky Barrier Diode was used. PMOS and NMOS transistor was selected. In order to use 5Spice Electronic Circuit Simulation Program, PMOS and NMOS transistors' properties should be defined. These properties are given in Table 1 and definitions and units of the parameters are given in Table 2.

NMOS 5SPICE MODEL		
LEVEL = 3		
+TOX = 7.9E-9	NSUB = 1E17	GAMMA = 0.5827871
+PHI = 0.7	VTO = 0.5445549	DELTA = 0
+UO = 436.256147	ETA = 0	THETA = 0.1749684
+KP = 2.055786E-4	VMAX = 8.309444E4	KAPPA = 0.2574081
+RSH = 0.0559398	NFS = 1E12	TPG = 1
+XJ = 3E-7	LD = 3.162278E-11	WD = 7.046724E-8
+CGDO = 2.82E-10	CGSO = 2.82E-10	CGB0 = 1E-10
+CJ = 1E-3	PB = 0.9758533	MJ = 0.3448504
+CJSW = 3.777852E-10	MJSW = 0.3508721	
PMOS 5SPICE MODEL		
LEVEL = 3		
+ TOX = 7.9E-9	NSUB = 1E17	GAMMA = 0.4083894
+ PHI = 0.7	VTO = -0.7140674	DELTA = 0
+ UO = 212.2319801	ETA = 9.999762E-4	THETA = 0.2020774
+ KP = 6.733755E-5	VMAX = 1.181551E5	KAPPA = 1.5
+ RSH = 30.0712458	NFS = 1E12	TPG = -1
+ XJ = 2E-7	LD = 5.000001E-13	WD = 1.249872E-7
+ CGDO = 3.09E-10	CGSO = 3.09E-10	CGB0 = 1E-10
+ CJ = 1.419508E-3	PB = 0.8152753	MJ = 0.5
+ CJSW = 4.813504E-10	MJSW = 0.5	

Table 1. 5Spice PMOS and NMOS Characteristic Data

LEVEL	Model type (1, 2, or 3)	
L	Channel length	meters
W	Channel width	meters
LD	Lateral diffusion length	meters
WD	Lateral diffusion width	meters
VTO	Zero-bias threshold voltage	Volts
KP	Transconductance	Amps/Volts ²
GAMMA	Bulk threshold parameter	Volts ^{1/2}
PHI	Surface potential	Volts
LAMBDA	Channel-length modulation	Volts ⁻¹
	(LEVEL = 1 or 2)	
RD	Drain ohmic resistance	Ohms
RS	Source ohmic resistance	Ohms
RG	Gate ohmic resistance	Ohms
RB	Bulk ohmic resistance	Ohms
RSH	Drain-source diffusion sheet resistance	Ohms/square
PB	Bulk p-n potential	Volts

CBD	Bulk-drain zero-bias p-n capacitance	Farads
CBS	Bulk-source zero-bias p-n capacitance	Farads
CJ	Bulk p-n zero-bias bottom capacitance/length	Farads/meters ²
CJSW	Bulk p-n zero-bias perimeter capacitance/length	Farads/meters
MJ	Bulk p-n bottom grading coefficient	
MJSW	Bulk p-n sidewall grading coefficient	
FC	Bulk p-n forward-bias capacitance coefficient	
CGSO	Gate-source overlap capacitance/channel width	Farads/meters
CGDO	Gate-drain overlap capacitance/channel width	Farads/meters
CGB0	Gate-bulk overlap capacitance/channel width	Farads/meters
NSUB	Substate doping density	1/centimeter ³
NSS	Surface-state density	1/centimeter ²
NFS	Fast surface-state density	1/centimeter ²
TOX	Oxide thickness	meters
TPG	Gate material type: + 1 = opposite of substrate, - 1 = same as substrate, 0 = aluminum	
XJ	Metallurgical junction depth	meters

UCRIT	Mobility degradation critical field (LEVEL = 2)	Volts/centimete r
UEXP	Mobility degradation exponent (LEVEL = 2)	
UTRA	(Not Used) mobility degradation transverse field coefficient	
VMAX	Maximum drift velocity	meters/seconds
NEFF	Channel charge coefficient (LEVEL = 2)	
XQC	Fraction of channel charge attributed to drain	
DELTA	Width effect on threshold	
THETA	Mobility modulation (LEVEL = 3)	Volts-1
ETA	Static feedback (LEVEL = 3)	
KAPPA	Saturation field factor (LEVEL=3)	
KF	Flicker noise coefficient	
AF	Flicker noise exponent	

Table 2. Spice Model parameters definitions and units of MOSFETS

In these simulations, PMOS transistor is used in order to deliver current into the heart and NMOS transistor is used in order to receive current from the heart. Simulations are made for unipolar and bipolar pacing modes.

Firstly, unipolar pacing mode was simulated. Series combination of PMOS transistor and diode were used in order to deliver pulse through electrode tissue impedance and series combination of NMOS transistor and diode were used in order to receive pulse. 1msec pulse was applied in the period of 1 sec. As a result of this simulation, at positive cycle, 4.1 volt

pulse level was observed and at negative cycle -4.4 volt pulse was observed on the electrode tissue impedance at 1msec. This result is seen in Figure 15a and 15b.




Figure 15a. Simulation result of TDC for unipolar pacing mode with positive pulse




Figure 15b. Simulation result of TDC for unipolar pacing mode with negative pulse

Bipolar pacing mode was simulated. 1msec pulse was applied in the period of 1 sec. As a result of this simulation, 3.6 volt pulse level observed on the electrode-tissue impedance at 1msec. This result is seen in Figure 16.




Figure 16. Simulation result of TDC for bipolar pacing mode

4.1.3 CAPACITOR SWITCH CIRCUIT (CSC)

Capacitor Switch Circuit was simulated by using 5Spice Electronic Circuit Simulation Program (Non-commercial Version, Richard P. Andresen, 2004). Capacitor was selected as $C=10\ \mu\text{F}$. Serial resistors selected as $27\ \text{k}\Omega$. In this simulation, Pulse level was selected as 5 volt. Pulse width of the source was selected as 999ms during charging period and source is turn off at 1msec during discharging period. During 1msec, capacitor discharged on $1\ \text{k}\Omega$ electrode tissue impedance when switching transistor was open. Result of CSC-parallel

capacitor- is sketched in Figure 17a and result of CSC-series capacitor- is sketched in Figure 17b.




Figure 17a. Simulation result of CSC-parallel capacitor- for bipolar pacing mode




Figure 17b. Simulation result of CSC-series capacitor- for bipolar pacing mode

4.2 PART II: INDUCED VOLTAGE SIGNAL MODEL

In these simulations, it is assumed that proposed device that contains IPG and wires is placed in MRI scanner. 5 cm length of IPG width and 35 cm length of wire are placed in parallel to electric field of the MRI scanner. It is stated in the Halise's Irak master thesis that 37.5 volt rms induced voltage on the electrode tissue impedance is calculated for 35 cm copper wire. This voltage value will be used as a signal source above simulations [26].

4.2.1 DIODE RESISTOR CIRCUIT (DRC)

In order to simulate this implementation, OrCAD PSpice 9.1 Demo (Cadence, 2655 Seely Avenue, San Jose, California 95134, USA) was used. In this simulation, Infineon Technologies BA595 pin diode model was used. 37.5 volt rms induced voltage at 64 MHz applied as a signal source. Resistance of the resistor was selected as 100 k Ω . Result of this simulation is given in Table 3.

As a result of this simulation, approximately 3.2 volt rms induced voltage was observed on the electrode (1 k Ω electrode tissue impedance). It is concluded that DRC decreases significantly voltage level at the electrode of the lead. Gradient field simulation was not performed because PIN diode cannot blocks low frequency current flow.

4.2.2 TRANSISTOR DIODE CIRCUIT (TDC)

In this simulation, 5Spice Electronic Circuit Simulation Program (Non-commercial Version, Richard P. Andresen, 2004) was used. 1N5711 Schottky Barrier Diode was used. Table 1 MOSFET Models were used for PMOS and NMOS transistor. 37.5 volt rms induced voltage at 64 MHz was applied as a signal source. As a result of this simulation, approximately 1 volt rms induced voltage was observed at the electrode impedance. 1 volt rms induced voltage was observed on the ring impedance. Significant voltage drop was

observed on the electrode tissue impedance in this simulation result. This result is seen in Table 3.

In addition to RF simulation, gradient field simulation was performed. In order to compare these results with RF induced voltage response of my designs, 37.5 volt rms induced voltage at 1 KHz applied as a signal source. The voltage was measured on electrode tissue impedance of the electrode and ring of the lead and the voltage value dropped significantly in this simulation. Its value was measured as 0.37 volt rms voltage and this result showed that this implementation block RF induced voltage effectively. This result is seen in Table 3.

4.2.3 CAPACITOR SWITCH CIRCUIT (CSC)

In the simulation of this implementation, 27 $K\Omega$ resistors are placed on the leads. 5Spice Electronic Circuit Simulation Program (Non-commercial Version, Richard P. Andresen, 2004) was used. 37.5 volt rms induced voltage at 64 MHz was applied as a signal source and CSC -parallel capacitance model- was simulated. As a result of this simulation, approximately 1.4 volt rms induced voltage was observed on the electrode tissue impedance and approximately 1.7 volt rms induced voltage was observed on the ring electrode tissue impedance. This result is seen in Table3.

Also gradient field simulation was performed. 37.5 volt rms induced voltage at 1 KHz applied as a signal source. Simulation result of induced voltages at the electrode tissue impedance and ring tissue impedance are seen in Figure 18a and 18b.




Figure 18a. Simulation result of CSC-parallel capacitor- on the electrode tissue impedance at 1 KHz




Figure 18b. Simulation result of CSC-parallel capacitor- on the ring tissue impedance at 1KHz

Induced voltage level and shape was different on the electrode and ring impedances. Reason of this situation is PMOS transistors' placement on the electrode wire. In fact, still measured induced voltage levels significantly decrease on both electrode and ring tissue impedances. Same simulations were repeated for CSC-series capacitor- implementation at 64 MHz and 1 KHz frequency. This result is seen in Table 3.

Design	Measured voltage on electrode-ring tissue impedance (Vrms)	
	64 MHz	1 KHz
DRC	3.2	-----
TDC	1	0.37
CSC –Parallel Capacitor	1.4(electrode)	0.64(electrode)
	1.7(ring)	2.1(ring)
CSC- Series Capacitor	1.56	1.3

Table 3. Simulation result of DRC, TDC and CSC voltage measurement on the electrode and ring impedances.

4.3 PART III: EFFICIENCY MODEL OF CSC

Efficiency equation of CSC was solved by using MATLAB (The MathWorks Inc., Natick, Massachusetts, USA) with respect to R_1 and C values. In Figure 19, the efficiency is plotted as a function of R_1 for $C= 10 \mu F$, $t_1=1 \text{ sec}$, $t_2=1 \text{ msec}$, $R_2=1 \text{ k}\Omega$.




Figure 19. The efficiency plot of CSC as a function of R_1

In Figure 20, the efficiency is plotted as a function of capacitance for $R_1=50 \text{ k}\Omega$, $t_1=1 \text{ sec}$,
 $t_2=1 \text{ msec}$, $R_2=1 \text{ k}\Omega$.




Figure 20. The efficiency plot of CSC as a function of capacitance

According to this solution, capacitance was selected as $C= 10 \mu\text{F}$ and resistance of a resistor was selected as $50 \text{ k}\Omega$. The efficiency was greater than 15 ($\eta \geq 15$) for these values. In order to block passage of RF current from gate of MOSFET to electrode and ring, $100 \text{ k}\Omega$ resistor is put on the control line of MOSFET transistor gate for both implementations. This resistor does not affect pulse process because there is no current on the MOSFET transistor gate. Thus, DC voltage level on the control line is not changed by putting this resistor.

In Figure 21, the efficiency is plotted as a function of t_1 for $C= 10 \mu\text{F}$, $R_1=50 \text{ k}\Omega$, $t_2=1 \text{ msec}$, $R_2=1 \text{ k}\Omega$.




Figure 21. The efficiency plot of CSC as a function of t_1

In Figure 22, the efficiency is plotted as a function of t_2 for $C=10\text{ }\mu\text{F}$, $R_1=50\text{ k}\Omega$, $t_1=1\text{ sec}$, $R_2=1\text{ k}\Omega$.




Figure 22. The efficiency plot of CSC as a function of t_2

These plots give idea about effect of charging and discharging duration on the efficiency of the system.

According to above calculations, specific resistance and capacitance values are determined.

In this design, modification is available when resistance distribute on the wire. Resistance of wire can be calculated as:

$$R = \frac{\rho L}{A} \quad (6)$$

where ρ is resistivity, L is length, A is the cross sectional area of the wire.

Based on the calculation above, the optimum resistance which is placed on the electrode wire is calculated as 25 k Ω . The same resistance is also placed on the ring wire. If it is assumed that length of the wire is 50 cm and copper is used as a material. Resistivity of copper is $\rho=1.724 \times 10^{-8} \Omega m$ at 20 C. By using this data, diameter of the copper wire is calculated as $d=6.63 \times 10^{-4} \text{ mm}$

In addition to this, if another element is used, for example carbon ($\rho=3.5 \times 10^{-5} \Omega m$), the new diameter of the wire will be $d=2.986 \times 10^{-2} \text{ mm}$. These two calculations show that a very thin wire can be used and also different materials, which are highly resistive, can be used in this design.

5. EXPERIMENTS

In this chapter, the experimental methods used to observe performance of the proposed designs are explained. Phantom heating experiments of wire with and without RF blocking elements for DRC and CSC are performed. The nerve stimulation experiment of CSC is also performed in the following part. The experimental method is described below.

5.1 HEATING MEASUREMENT EXPERIMENTS

In the first section, heating experiment setups of wire with and without RF blocking elements are explained. In the second and third sections, heating experiment setups of DRC and CSC are explained.

5.1.1 COMPARISON OF WIRE HEATING WITH AND WITHOUT RF BLOCKING ELEMENTS (RFBE)

Experiment was performed to observe heating response of wire with and without RF blocking elements. In this experiment, Fast SPGR sequence was used with following imaging parameters: Body coil (diameter: 60 cm), matrix 256x256, NEX: 4, TE: min, TR: 19 msec, flip angle: 90, bandwidth: 62.5 KHz, FOV: 48, Slice thickness: 20 mm.

The Scanner software estimated the average SAR as 1.26 W/Kg, and peak SAR as 2.51 W/Kg. It should be noted that, while scanner estimates these numbers, it assumes that a person as weighs 30 Kg., and therefore, the estimation method used in the scanner may fail to display the correct number.

A semi cylindrical phantom (19.5 cm diameter, 50 cm length, and see Figure 23) was constructed that enabled accurate and reproducible placement of wire and temperature measuring probes in a gel to measure temperature increase.

Semi Cylindrical Phantom Model




Figure 23. Semi cylindrical phantom model.

The phantom consisted of a gel. This gel consisted of 4 lt. water, 1.6 kg gel flour and 50 gr. salt without iodine.

Copper hookup wire with 0.5 mm diameter was straightened and cut into 30 cm lengths. Three cases were tested: wire without any RF blocking element on it, wire with 3.9 k Ω resistor placed on the distal electrode of the wire and wire with PIN diode MA4P7452F-1072T (M/A-COM, Pawtucket Blvd., Lowell, USA) placed on the distal electrode of the wire.

The gel phantom was placed longitudinally in the scanner, with the wire placed longitudinally in it. In order to obtain high electrical field strength in the gel phantom, it was placed within 0 to 1 cm away from main body coil of the scanner. In addition to this, the lead was placed within 3 to 5 cm away from phantom edge that is close to body coil of the scanner. According to this adjustment, bare sides of lead were placed inside the gel, and remaining part was placed on the surface of the gel.




Figure 34. Lead placement in the gel




Figure 35. Phantom and lead placement in the scanner




Figure 26. Lead model without RFBE used in the heating measurement




Figure 27. Lead model with RFBE used in the heating measurement

5.1.2 DIODE RESISTOR CIRCUIT (DRC)

Experiment was performed to observe heating response of DRC. In this experiment, Fast SPGR sequence was used with following imaging parameters: Body coil (diameter: 60 cm), matrix 256x256, NEX: 4, TE: min, TR: 19 msec, flip angle: 90, bandwidth: 62.5 KHz, FOV: 48, Slice thickness: 20 mm.

The Scanner software estimated the average SAR as 1.26 W/Kg, and peak SAR as 2.51 W/Kg. The phantom is semi cylinder with a 10 cm radius and a 50 cm length. This gel phantom setup is seen in Figure 28a and 28b.



Figure 28a. Gel phantom setup with DRC



Figure 28b. Gel phantom setup with DRC

5.1.3 CAPACITOR SWITCH CIRCUIT (CSC)

With the aim of testing the CSC- parallel capacitor- circuit, a circuit that emulates implantable pace generator (IPG) was designed. The simple pulse circuit contained a standard 9 Volt battery, and a 16F84A 18-pin Enhanced FLASH/EEPROM 8-Bit microcontroller. 16F84A microcontroller was programmed to generate 1msec pulse with a period of 1 sec. 9 Volt battery and 16f84A microcontroller were put on the same board and connected to each other. This board was put into a waterproof plastic box with the dimension of 9.5x5x2.5 (cm). Approximately 29 cm copper wires (0.5 mm diameter) were connected to this box .The waterproof plastic box was covered with copper tape and this conductive coating was connected to ground wire.

Different than IPG box of this design, another mini box design was made in order to place CSC on the lead of the IPG. In this mini box, 27 k Ω resistor was used. 500 k Ω resistor connected to gate of PMOS transistor due to obtain reliable gate isolation for RF signal during MRI examination. 10 μ F tantalum capacitor and BS250 PMOS transistor were placed in the circuit. CSC (parallel capacitor) was built by using these components.

After mini box that contain CSC was added at the end of the lead, total length of wire was approximately 35 cm and my safe implantable circuit design was created simply. In order to compare the safety performance of CSC implementation, an IPG model without CSC circuit was created. Two designs are seen in Figure 29a and Figure 29b.




Figure 29a. IPG model without CSC




Figure 29b. IPG model with CSC

These two circuits were put into approximate rectangular plastic phantom which dimension is 51x14.7x11 (cm) (Figure 30a and 30b).




Figure 30a. Rectangular plastic phantom side 1



Figure 30b. Rectangular plastic phantom side 2

These gel phantom setups are seen in Figure 31a and Figure 31b.




Figure 31a. Experiment setup without CSC




Figure 31b. Experiment setup with CSC

In order to obtain worst tissue heating response, gel was used. 5 lt. water, 2 kg gel dust and 50 gr. salt without iodine put into plastic case and then mixed each other in the case. 3 hours later, IPG with CSC was placed in one case and standard IPG without CSC was placed in another case. By using these phantoms, experiments were performed.

Heat measurement experiment was performed for CSC in GE Signa 1.5T MR scanner, MRI Unit in Gazi Üniversitesi Tıp Fakültesi (Gazi Hastanesi Beşevler / Ankara). In this experiment, SPGR sequence was used with following imaging parameters: Body coil, matrix 256x256, NEX: 4, TE: 4.0 msec, TR: 19 msec, flip angle: 17, bandwidth: 16.53 KHz, total scan time: 10.59 sec

Scanner software estimated average SAR as 1.2328 W/Kg, and peak SAR as 2.4656 W/Kg.

5.2 NERVE STIMULATION EXPERIMENT

In this chapter, nerve stimulation experiment setup of CSC is given.

5.2.1 CAPACITOR SWITCH CIRCUIT (CSC)

In order to test gradient field effect to see if implanted devices increases the possibility of nerve stimulation in a MRI scanner during a routine MRI exam, a frog leg experiment was performed. Both IPG with and without CSC were tested with the same experimental set-up to show that new design improves nerve stimulation safety aspect of the implanted devices.

The frog leg experiments were conducted in a GE Signa 1.5T MR scanner in Gazi University School of Medicine, Ankara, Turkey¹. In this experiment, the gastrocnemius nerve stimulation that can easily be seen by the naked eye was used. The preparation of the Sciatic Nerve for this experiment is explained below:

The frog dorsal was turned side up. The lower part of the thoracic cavity was cut (a U shaped area with the Urostyle down the center) taking care not to damage the cavity, since this would result in severing the Sciatic nerve. The Sciatic nerve was tied off with a length of thread just as it emerges from the spinal column [27]. It is seen in Figure 32.




Figure 32. The Sciatic Nerve bundle where it leaves the vertebral column [27].

¹ Dr. Sevin Guney helped in preparation of the frog experiment.

The Sciatic nerve was exposed as it runs through the hip joint using blunt glass tools (i.e. two glass rods drawn out and fire-polished to a blunt 1 mm tip). Sciatic nerve runs through the thigh musculature. It is seen in Figure 33.




Figure 33. Sciatic Nerve in thigh musculature [27]

Finally, the Sciatic nerve was free from the spinal column down to the point it enters the Gastrocnemius. The nerve muscle preparation was placed in a ringer solution; a solution of recently boiled distilled water that contains 8.6 gram sodium chloride, 0.3 gram potassium chloride, and 0.33 gram calcium chloride per lt. Ringer solutions containing the chlorides of sodium, potassium, calcium and magnesium, in order to obtain a suitable physiological saline solution which would keep the leg of the frog alive outside of the body. This solution keeps the prepared frog leg alive for more than two hours. It is seen in Figure 34.



Figure 34. Nerve muscle preparation of the frog leg

Next, the threshold value of the nerve stimulation is needed to be found using the pulse generator. These measurements provided an overall understanding about the threshold level of the nerve stimulation for frog leg. The frequency was adjusted to 1 Hz, the pulse width to 1 msec, and the voltage values to 5 volt. Stimulation of the frog leg was observed when the 5 volt was applied. Then this voltage values were decreased until stimulation was not observed on the frog leg. From this measurement, it was concluded that the threshold level of the gastrocnemius nerve was approximately above 0.5 Volt. Experiment setup was prepared by putting IPG emulator into plastic case. This is seen in Figure 35.



Figure 35. Nerve stimulation experiment setup

Then this setup was filled with ringer solution and frog leg was put into this setup. Sciatic nerve of frog leg was connected to lead of the emulator. Measurements were made with and without CSC. For both experiments, all conditions were kept the same.

6. RESULTS

In this chapter, results of heating measurement experiments and nerve stimulation experiments are given.

6.1 HEAT MEASUREMENT EXPERIMENT RESULTS

In this part, in the first section, the results of wire heating with and without RF blocking elements and in the second, third sections, the results of DRC and CSC heating experiments are given.

6.1.1 COMPARISON OF WIRE HEATING WITH AND WITHOUT RF BLOCKING ELEMENTS (RFBE)

The RF heating at the wire electrodes was measured for three different configurations each of 20 cm and 30 cm wires. The temperature probe 1 was placed at the surface of the distal electrode of the wires. The temperature probe 2 was placed at the surface of the proximal electrode of the wires.

Figures 36-38 shows RF heating temperature measurements of 20 cm wires. Figure 36 shows heating of a 20 cm wire without any RF blocking element. Figure 37 shows heating of the wire with a 3.9 k Ω resistor placed on the distal. Figure 38 shows heating with PIN diode MA4P7452F-1072T placed on the distal electrode of the wire.

Figures. 39-41 show the RF heating temperature measurement of 30 cm wires. Figure 39 shows heating of a 30 cm wire without any RF blocking element. Figure 40 shows RF heating with 3.9 k Ω resistor placed on the distal electrode of the wire. Figure 41 shows RF heating with PIN diode MA4P7452F-1072T placed on the distal electrode of the wire.




Figure 36. Temperature rise versus time at the distal and proximal electrode of a 20 cm wire without RF Blocking element




Figure 37. Temperature rise versus time at the distal and proximal electrode of the 20 cm wire 3.9 k Ω resistance placed on the distal electrode of the wire




Figure 38. Temperature rise versus time at the distal and proximal electrode of a 20 cm wire with PIN diode MA4P7452F-1072T placed on the distal electrode of the wire




Figure 39. Temperature rise versus time at the distal and proximal electrode of a 30 cm wire without RF blocking element




Figure 40. Temperature rise versus time at the distal and proximal electrode of the 30 cm wire 3.9 kΩ resistance placed on the distal electrode of the wire




Figure 41. Temperature rise versus time at the distal and proximal electrode of a 30 cm wire with PIN diode MA4P7452F-1072T placed on the distal electrode of the wire

6.1.2 DIODE RESISTOR CIRCUIT (DRC)

In the heating experiment of Diode_Resistor Circuit (DRC), the temperature was recorded using ReFlex-4 (Neoptix Inc., 1415, blvd Charest Quest, Suite 220, Quebec, QC, G1N 4N7, Canada) TMI Signal Conditioner and fiberoptic temperature sensor probes (Neoptix Inc., 1415, blvd Charest Quest, Suite 220, Quebec, QC, G1N 4N7, Canada). Phantom placed longitudinally in the MRI scanner. Its position is close to body coil. Temperature probe situated at a constant position at the end of the electrode, and placed longitudinally in the phantom. In this measurement, temperature increase approximately 3 centigrade in 6 minutes. This measurement plot is given in Figure 42.




Figure 42. Heating experiment result of IPG model with DRC

6.1.3 CAPACITOR SWITCH CIRCUIT (CSC)

The measurements were made for IPG with and without CSC in the same conditions. The temperature was recorded using FISO TMI Signal Conditioner and FOT-M fiberoptic temperature probes (FISO Technologies, Ste. Foy, Quebec, Canada). Phantom placed longitudinally in the MRI scanner. Its position is close to body coil and this position is same

for both circuit measurements. Temperature probe situated at a constant position on the surface of the electrode, and placed longitudinally in the phantom. In the first measurement, IPG without CSC placed in MRI scanner. In this measurement, temperature increase was more than 15 centigrade in 10 minutes. This measurement plot is given in Figure 43.




Figure 49. Heating experiment result of IPG model without CSC

In the second measurement, IPG with CSC placed in MRI scanner. In this measurement, temperature increase was approximately 0.5 centigrade in the same period of time. This measurement plot is given in Figure 44.




Figure 44. Heating experiment result of IPG model with CSC

In both plots, MRI sequence was started at 125th sec and finished at 750th sec after measurement was initialized.

6.2 NERVE STIMULATION EXPERIMENT RESULTS

6.2.1 CAPACITOR SWITCH CIRCUIT (CSC)

In the experiments, spin-echo single shot echo-planar sequence was used. For each of the designs, 5 different FOV values (5, 15, 31, 45 and 48 cm) were used. Note that changing FOV increases amplitude of the readout gradients and therefore increase duration of the ramp time (in GE scanners, the slope of the gradient waveform is kept constant). Therefore, low FOV values should result in longer exposure to the dB/dt. Other pulse sequence parameters were: matrix 256x256, TE: 95.4, TR: minimum, slice thickness: 7 mm, bandwidth: 20.63 KHz.

By using these properties, two experiments were conducted. In the first experiment, stimulation was observed on the frog leg for the IPG without CSC in each FOV value. On the other hand, in the second experiment, stimulation creation was not observed for the IPG with CSC in each FOV value. This experiment shows that CSC can provide significant opportunity to prevent nerve stimulation for the patients who carry the stimulators.

7. DISCUSSIONS

In this thesis, a new method is proposed to make MRI safe implanted devices. This method uses active circuits on the lead in order to prevent induced currents on the wires. Here, three implementations of this novel technique are proposed. Each of them has different properties and is preferred in different implementations. By simulations and heating experiments, it is observed that the proposed technique was successful in significantly reducing the induced currents.

PIN diodes' are very suitable for the reduction of RF induced current. Although they behave like a normal diode when a low frequency pacing signal is applied, they act as a resistor at radio frequencies where resistance is a function of applied DC current. When no voltage or negative voltage is applied to the diode, the RF resistance is in the order of several kilo ohms. Also the circuit can be miniaturized and easily incorporated into flexible leads without a significant effort and cost.

When the stimulator is exposed to radio frequency electromagnetic radiation, induced RF voltage on the lead would not cause excessive current on the electrode because of the diode and resistor in parallel in my design.

My circuit uses a large series capacitor and resistor in parallel with the diode. (Standard stimulator circuits use a similar series capacitor.) The series capacitor blocks the DC current flow to the target body part in a fault condition. After the pulse, some charge accumulates in the body, which is discharged by the resistor. The value of this parallel resistor must be chosen properly. If resistance is too high, the capacitor cannot be discharged during one pacing cycle and may adversely affect the pacing capability of the design. However, if the resistance is too low, some of the RF current may flow on the resistor and weaken the RF blocking capability of this circuit.

Modern pacemakers have different and various types of pacing pulse techniques. According to patient cardiac condition and illness type, unipolar, bipolar or different pacing pulse modes can be applied.

According to unipolar pacing mode, the pacing pulse current completes the loop from the single electrode and the case of the pacemaker, which is used as ground. This unipolar pacing pulse can be applied both directional. During a pacing pulse, the PIN diode conducts the pacing current after the pacing pulse level passes diode's turn-on voltage (typically 0.6-0.7V). A loss of 0.6V will be observed on the pacing voltage. For instance, if it is desired that 5V pacing pulse applies on the $1\text{k}\Omega$ electrode-tissue impedance, 5.6V pacing pulse should be applied in order to achieve this pacing pulse level on the electrode-tissue impedance. At this time, total power consumption is 28 mW where source voltage 5.6V and current 5 mA. Power loss due to pin diode is 3 mW where pin diode threshold voltage 0.6V and current 5 mA. The ratio between power loss and total power is 10.71%. This ratio decreases when the source voltage increases. On the other hand, the ratio increases when the source voltage decreases. However, the total power consumption decrease with this ration and the power loss loses its weight becoming smaller.

Manufacturing process of the Diode Resistor Circuit (DRC) is rather straightforward. No tuning is necessary. Very small sized diodes and resistors are produced in today technology. Therefore, this circuit can be built in very small area without many difficulties and without changing wire design.

RF heating experiments has showed the performance of the Diode Resistor Circuit. These experiments using 20 cm and 30 cm wire without any RF blocking elements show that heating is significant (14.6°C temperature increase in 290 seconds for 20 cm wire and 21.6°C temperature increase in 240 seconds for 30 cm wire) at the distal and proximal

electrodes of the wires. The temperature increases with time as soon as the series of RF pulses begins and decreases quickly after it stops.

The results in RF heating experiments of 20 cm and 30 cm wire with a 3.9 k Ω resistor indicate that the heating decreased significantly (1°C temperature increase in 240 seconds for 20 cm wire and 2.1°C temperature increase in 350 seconds for 30 cm wire) at the distal electrode. However, heating is still important (4.8°C temperature increase in 240 seconds for 20 cm wire and 14.7°C temperature increase in 350 seconds for 30 cm wire) at the proximal electrodes of the wire.

The results in these experiments of 20 cm and 30 cm wire with PIN diode MA4P7452F-1072T indicate that the heating decreased significantly (1.1°C temperature increase in 305 seconds for 20 cm wire and 4.7°C temperature increase in 510 seconds for 30 cm wire) at the distal electrode. Similar to measurement of wire with resistor, heating at the proximal electrode is still important (5°C temperature increase in 305 seconds for 20 cm wire and 21.8°C temperature increase in 510 seconds for 30 cm wire).

The second implementation, Transistor Diode Circuit (TDC) is advance model of DRC implementation model. Therefore, very similar properties are valid for this model. High impedance circuit elements are used at radio frequencies similar to DRC. In addition to this, it provides a contribution to the safety condition of stimulators; because the gradient induced current can be decreased using this implementation.

The third implementation model, Capacitor Switch Circuit (CSC) is suitable for the bipolar pacing technique. Although miniaturization of this design can be more challenging compared to the other two implementations, this design provides safety for both gradient and RF induced currents. This design can be implemented by using series and parallel capacitors. While a parallel capacitor design may suffer from an electrode polarization problem, a series capacitor design needs additional control lines as it has been implemented.

In addition, in the nerve stimulation experiment of CSC, the exact threshold level of the gastrocnemius nerve stimulation was not determined. Because some parameters such as gradient field amplitude and duration were not measured during these experiments the nerve stimulation experiments were not enough to judge if there is a significant nerve stimulation risk associated with implants. However, if it is assumed that there is a risk; the CSC design was shown to be reducing this effect.

Although my implementations are suitable for the most active implant designs, some active designs may require additional circuits. Implementation of these circuits is trivial given the above designs. For example, implantable cardioverter defibrillators (ICDs) require that circuit elements sustain large voltages and currents. Cardiac pacemakers will need an additional circuit for sensing ECG signals.

Deep Brain Stimulation is another basic area for stimulators. It delivers electrical stimulation signals to targeted areas in the brain. DBS is used for Parkinson's disease, Essential Tremors, and Dystonia. DBS has three parts. These are lead, IPG and extension cable. DBS leads may have four or more electrodes. Extension is the connection point between IPG and DBS lead. This proposed active circuit may be placed outside the brain just on the skull where the DBS lead is connected to the extension cable. The implementation of the design can be varied.

8. CONCLUSION

In this thesis, active circuits were proposed in order to block induced currents on the implant leads. Three different implementation of this design were shown. The circuit designs are novel over other possible designs that can be found in the literature. These designs are very effective in blocking induced currents which may be due to radio frequencies and gradient fields as shown by computer simulations and phantom and animal experiments. These models can be implemented in a very small area without much difficulty. In my implementations, no tuning is required. It was seen that the new techniques are effective in making MRI safe and implantable devices. The benefits and problems of each technique were discussed in this thesis. It is believed that by using this or similar techniques, patients with implants can be examined safely in MRI scanners.

9. REFERENCES

- [1] Medscape, <http://www.medscape.com/viewarticle/503383>, Medscape CRM News 2005.
- [2] Pennsylvania Health Care Cost Containment Council (PHC4), <http://www.phc4.org/reports/fyi/fyi27.htm>, 2007.
- [3] Kalin R, Stanton MS. Current clinical issues for MRI scanning of pacemaker and defibrillator patients. *Pacing Clin Electrophysiol.* 2005;28:326–328.
- [4] C. D. Smith, A. V. Kildishev, J. A. Nyenhuis, K. S. Foster, and J. D. Bourland, “Interactions of magnetic resonance imaging radio frequency magnetic fields with elongated medical implants”, *J Applied Physics*, vol. 87, no. 9, 2000.
- [5] T. Sommer, C. Vahlhaus, G. Lauck, A. V. Smekal, M. Reinke, U. Hofer, W. Block, F. Träber, C. Schneider, J. Gieseke, W. Jung, H. Schild, “MR Imaging and Cardiac Pacemakers: In Vitro Evaluation and in Vivo Studies in 51 Patients at 0.5 T”, *Radiology*, vol. 215, no. 3, 2000.
- [6] R. C. Luechinger, “Safety aspects of cardiac pacemakers in magnetic resonance imaging,” Ph.D. dissertation, Dept. Naturwissenschaften, Swiss Federal Inst. Technol., Zurich, Switzerland, 2002.
- [7] FDA Center for Devices and Radiological Health. Manufacturer and User Facility Device Experience Database—(MAUDE). [Online]. Available: http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfMAUDE/Detail.CFM?MDRFOI_ID=315105
- [8] FDA Center for Devices and Radiological Health. Manufacturer and User Facility Device Experience Database—(MAUDE). [Online]. Available: http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfMAUDE/Detail.CFM?MDRFOI_ID=474005
- [9] A. R. Rezai, D. Finelli, J. A. Nyenhuis, G. Hrdlicka, J. Tkach, A. Sharan, P. Rugieri, P. H. Stypulkowski, and F. G. Shellock, “Neurostimulation systems for deep brain stimulation: In vitro evaluation of magnetic resonance imaging-related heating at 1.5 T,” *J. Magn. Reson. Imaging*, vol. 15, no. 3, pp. 241–250, Mar. 2002.
- [10] W. Kainz, G. Neubauer, R. Uberbacher, F. Aleschand, and D. D. Chan, “Temperature measurement on neurological pulse generators during MR scans,” *Biomed. Eng. Online*, vol. 1, no. 2, Sep. 12, 2002. [Online]. Available: <http://www.biomedical-engineering-online.com/content/1/1/2>
- [11] S. R. Benbadis, J. Nyenhuis, W. O. Tatum, IV, F. R. Murtagh, M. Gieron, and F. L. Vale, “MRI of the brain is safe in patients implanted with the vagus nerve stimulator,” *Seizure*, vol. 10, no. 7, pp. 512–515, Oct. 2001.
- [12] M. F. Dempsey and B. Condon, “Thermal injuries associated with MRI,” *Clin. Radiol.*, vol. 56, no. 6, pp. 457–465, Jun. 2001.

- [13] U.S. Department of Health and Human Services, Food and Drug Administration, Center for Devices and Radiological Health, Guidance for the submission of premarket notifications for magnetic resonance diagnostic devices US DHHS FDA, Rockville, MD, 1998.
- [14] Yeung CJ, Susil RC, Atalar E. "RF safety of wires in interventional MRI: Using a safety index". *Magn Reson Med* 2002; 47:187-193.
- [15] Susil RC et al., "Multifunctional Interventional Devices for MRI: A Combined Electrophysiology/MRI Catheter", *MRM* 47:594-600 (2002).
- [16] Ladd ME et. al., "Reduction of Resonant RF Heating in Intravascular Catheters Using Coaxial Choke"s, *MRM* 43:615-619 (2000).
- [17] Thompson et al, "Implantable medical device incorporating integrated circuit notch filters", US Patent # 6,539,253, 2003
- [18] Nappholz et. al, "Rate-responsive pacemaker with minute volume determination and EMI protection", US Patent 5,817,136, 1998
- [19] Tsitlik J.E et. al., "ECG Amplifier and Cardiac Pacemaker for use during Magnetic Resonance Imaging", US Patent # 5,217,010, 1993.
- [20] Greatbatch W, Miller V, Shellock FG. "Magnetic resonance safety testing of a newly-developed, fiber-optic cardiac pacing lead". *J Magn Reson Imaging* 2002; 16:97–103.
- [21] Reilly JP, Diamant AM. "Theoretical evaluation of peripheral nerve stimulation during MRI with an implanted spinal fusion stimulator". *Magn Reson Imaging*. 1997;15(10):1145-56.
- [22] Medtronic, <http://www.medtronic.com/crm/kappa900/specs.html> , Medtronic, Inc. 2007.
- [23] F. Silveira and D. Flandre, *Low Power Analog CMOS for Cardiac Pacemaker- Design and Optimization in Bulk and SOI Technologies*, Kluwer Academic Publishers, Boston, 2004.
- [24] C.E. Hayes, "Radio frequency field coil for NMR", US Pat. 4692705 - Filed Dec 23, 1983 - General Electric Company.
- [25] K. Brandley, "The Technology: The Anatomy of a spinal Cord and Nerve Root Stimulator: The Lead and The Power Source", *Pain Medicine*, vol. 7, number S1, 2006.
- [26] Halise Irak, "Modeling RF Heating of Active Implantable Medical Devices During MRI Using Safety Index", Dissertation for the Degree of Master of Science, Bilkent University, Electrical and Electronics Engineering Department, 2007.
- [27] Neuromuscular Physiology, <http://mipwww.life.uiuc.edu/403%20Labs/Lab%20%2311-NeuroMusc/NeuroMusc40305.pdf>, 06/04/2004.